

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park
18 On Lai Street, Shatin
NT, Hong Kong
Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>



Report No. : 01656
Application No. : HP01393

Issue Date : 9 Jul 2026

TEST REPORT

Applicant : Cinotech Consultants Limited
Rm 1710, Technology Park, 18 On Lai Street,
Shatin, New Territories, Hong Kong.

Sample Description : Four (4) submitted sample(s) stated to be Wastewater.

Project No. : MA25111

Sampling date : 2 Jul 2026

Sample Status Upon Receipt : Room Temperature

Date Received : 2 Jul 2026

Test Period : 3 Jul 2026 to 3 Jul 2026

Test Requested : Total Suspended Solids (TSS)

Test Method : Method SOP001 reference to Standard Methods for the Examination of Water and Wastewater, 23rd Ed, 2017- APHA, WWA, WPCF; Part 2540 D.

Test Result : Refer to the test result(s) on page 2.

Remark : 1. Information of the sample description (except sample status upon receipt) provided by the Applicant.
2. The result(s) relate only to the items tested.
3. The result(s) apply to the sample as received, where HPCT has not been responsible for the sampling stage.

For and on behalf of
HIGH PRECISION CHEMICAL TESTING LIMITED

Chan Hon Fai
Laboratory Director

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01656
Application No. : HP01393

Issue Date : 9 Jul 2026

TEST REPORT

Test Result : Total Suspended Solids (TSS)

Client sample ID	Sampling date	Laboratory ID	Result (mg/L)	Uncertainty (mg/L)
C1_1	2 Jul 2026	HP01393-01	1.8	± 0.1
C1_2	2 Jul 2026	HP01393-02	1.7	± 0.1
M1_1	2 Jul 2026	HP01393-03	2.4	± 0.2
M1_2	2 Jul 2026	HP01393-04	2.4	± 0.2

Note : 1. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%.
2. The reporting limit for Total Suspended Solids (TSS) is 1.0 mg/L.
3. mg/L denotes milligram per liter.

- End of report -

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01656QC

Issue Date : 10 Aug 2026

Application No. : HP01393

QC REPORT

Applicant : Cinotech Consultants Limited
Rm 1710, Technology Park, 18 On Lai Street,
Shatin, New Territories, Hong Kong.

Sample Description : Four (4) submitted sample(s) stated to be Wastewater.

Project No. : MA25111

Sampling date : 2 Jul 2026

Sample Status Upon Receipt : Room Temperature

Date Received : 2 Jul 2026

Test Period : 3 Jul 2026 to 3 Jul 2026

Test Requested : Total Suspended Solids (TSS)

Test Method : Method SOP001 reference to Standard Methods for the Examination of Water and Wastewater, 23rd Ed, 2017- APHA, WWA, WPCF; Part 2540 D.

QC Record : Refer to the record(s) on page 2.

Remark : Information of the sample description (except sample status upon receipt) provided by the Applicant.

For and on behalf of
HIGH PRECISION CHEMICAL TESTING LIMITED

Chan Hon Fai
Laboratory Director

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01656QC

Issue Date : 10 Aug 2026

Application No. : HP01393

QC REPORT

	Description	Reported Result (mg/L)	Acceptance criteria	Conclusion
Quality control	Method Blank	< 0.20	< 0.2 mg/L	Pass
	Quality control Sample	19.8	17 to 23 mg/L	Pass
	Laboratory Fortified Blank	98.7	85 to 115 mg/L	Pass
	Duplicate Sample T1	1.7	% Difference within $\pm 15\%$	(%Diff = 4.5 %) Pass
	Duplicate Sample T2	1.8		

Note : 1. Perform method blank, quality control and laboratory fortified duplicate at a frequency of 1 in 20 samples or less.
2. mg/L denotes milligram per liter.
3. < denotes less than.
4. % denotes percent.

- End of report -

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park
18 On Lai Street, Shatin
NT, Hong Kong
Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>



Report No. : 01657
Application No. : HP01401

Issue Date : 10 Jul 2026

TEST REPORT

Applicant : Cinotech Consultants Limited
Rm 1710, Technology Park, 18 On Lai Street,
Shatin, New Territories, Hong Kong.

Sample Description : Four (4) submitted sample(s) stated to be Wastewater.

Project No. : MA25111

Sampling date : 3 Jul 2026

Sample Status Upon Receipt : Room Temperature

Date Received : 3 Jul 2026

Test Period : 6 Jul 2026 to 6 Jul 2026

Test Requested : Total Suspended Solids (TSS)

Test Method : Method SOP001 reference to Standard Methods for the Examination of Water and Wastewater, 23rd Ed, 2017- APHA, WWA, WPCF; Part 2540 D.

Test Result : Refer to the test result(s) on page 2.

Remark : 1. Information of the sample description (except sample status upon receipt) provided by the Applicant.
2. The result(s) relate only to the items tested.
3. The result(s) apply to the sample as received, where HPCT has not been responsible for the sampling stage.

For and on behalf of
HIGH PRECISION CHEMICAL TESTING LIMITED

Chan Hon Fai
Laboratory Director

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01657
Application No. : HP01401

Issue Date : 10 Jul 2026

TEST REPORT

Test Result : Total Suspended Solids (TSS)

Client sample ID	Sampling date	Laboratory ID	Result (mg/L)	Uncertainty (mg/L)
C1_1	4 Jul 2026	HP01401-01	2.8	± 0.2
C1_2	4 Jul 2026	HP01401-02	1.8	± 0.1
M1_1	4 Jul 2026	HP01401-03	2.7	± 0.2
M1_2	4 Jul 2026	HP01401-04	2.0	± 0.1

Note : 1. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%.
2. The reporting limit for Total Suspended Solids (TSS) is 1.0 mg/L.
3. mg/L denotes milligram per liter.

- End of report -

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park
18 On Lai Street, Shatin
NT, Hong Kong
Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>



Report No. : 01657QC
Application No. : HP01401

Issue Date : 10 Aug 2026

QC REPORT

Applicant : Cinotech Consultants Limited
Rm 1710, Technology Park, 18 On Lai Street,
Shatin, New Territories, Hong Kong.

Sample Description : Four (4) submitted sample(s) stated to be Wastewater.

Project No. : MA25111

Sampling date : 3 Jul 2026

Sample Status Upon Receipt : Room Temperature

Date Received : 3 Jul 2026

Test Period : 6 Jul 2026 to 6 Jul 2026

Test Requested : Total Suspended Solids (TSS)

Test Method : Method SOP001 reference to Standard Methods for the Examination of Water and Wastewater, 23rd Ed, 2017- APHA, WWA, WPCF; Part 2540 D.

QC Record : Refer to the record(s) on page 2.

Remark : Information of the sample description (except sample status upon receipt) provided by the Applicant.

For and on behalf of
HIGH PRECISION CHEMICAL TESTING LIMITED

Chan Hon Fai
Laboratory Director

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01657QC

Issue Date : 10 Aug 2026

Application No. : HP01401

QC REPORT

	Description	Reported Result (mg/L)	Acceptance criteria	Conclusion
Quality control	Method Blank	< 0.20	< 0.2 mg/L	Pass
	Quality control Sample	19.6	17 to 23 mg/L	Pass
	Laboratory Fortified Blank	98.3	85 to 115 mg/L	Pass
	Duplicate Sample T1	2.8	% Difference within $\pm 15\%$	(%Diff = 5.5 %) Pass
	Duplicate Sample T2	3.0		

Note : 1. Perform method blank, quality control and laboratory fortified duplicate at a frequency of 1 in 20 samples or less.
2. mg/L denotes milligram per liter.
3. < denotes less than.
4. % denotes percent.

- End of report -

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park
18 On Lai Street, Shatin
NT, Hong Kong
Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>



Report No. : 01658
Application No. : HP01402

Issue Date : 13 Jul 2026

TEST REPORT

Applicant : Cinotech Consultants Limited
Rm 1710, Technology Park, 18 On Lai Street,
Shatin, New Territories, Hong Kong.

Sample Description : Four (4) submitted sample(s) stated to be Wastewater.

Project No. : MA25111

Sampling date : 6 Jul 2026

Sample Status Upon Receipt : Room Temperature

Date Received : 6 Jul 2026

Test Period : 7 Jul 2026 to 7 Jul 2026

Test Requested : Total Suspended Solids (TSS)

Test Method : Method SOP001 reference to Standard Methods for the Examination of Water and Wastewater, 23rd Ed, 2017- APHA, WWA, WPCF; Part 2540 D.

Test Result : Refer to the test result(s) on page 2.

Remark : 1. Information of the sample description (except sample status upon receipt) provided by the Applicant.
2. The result(s) relate only to the items tested.
3. The result(s) apply to the sample as received, where HPCT has not been responsible for the sampling stage.

For and on behalf of
HIGH PRECISION CHEMICAL TESTING LIMITED

Chan Hon Fai
Laboratory Director

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01658
Application No. : HP01402

Issue Date : 13 Jul 2026

TEST REPORT

Test Result : Total Suspended Solids (TSS)

Client sample ID	Sampling date	Laboratory ID	Result (mg/L)	Uncertainty (mg/L)
C1_1	6 Jul 2026	HP01402-01	3.3	± 0.2
C1_2	6 Jul 2026	HP01402-02	4.3	± 0.3
M1_1	6 Jul 2026	HP01402-03	77.5	± 5.7
M1_2	6 Jul 2026	HP01402-04	81.5	± 5.9

Note : 1. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%.
2. The reporting limit for Total Suspended Solids (TSS) is 1.0 mg/L.
3. mg/L denotes milligram per liter.

- End of report -

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>



Report No. : 01658QC

Issue Date : 10 Aug 2026

Application No. : HP01402

QC REPORT

Applicant : Cinotech Consultants Limited
Rm 1710, Technology Park, 18 On Lai Street,
Shatin, New Territories, Hong Kong.

Sample Description : Four (4) submitted sample(s) stated to be Wastewater.

Project No. : MA25111

Sampling date : 6 Jul 2026

Sample Status Upon Receipt : Room Temperature

Date Received : 6 Jul 2026

Test Period : 7 Jul 2026 to 7 Jul 2026

Test Requested : Total Suspended Solids (TSS)

Test Method : Method SOP001 reference to Standard Methods for the Examination of Water and Wastewater, 23rd Ed, 2017- APHA, WWA, WPCF; Part 2540 D.

QC Record : Refer to the record(s) on page 2.

Remark : Information of the sample description (except sample status upon receipt) provided by the Applicant.

For and on behalf of
HIGH PRECISION CHEMICAL TESTING LIMITED

Chan Hon Fai
Laboratory Director

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01658QC

Issue Date : 10 Aug 2026

Application No. : HP01402

QC REPORT

	Description	Reported Result (mg/L)	Acceptance criteria	Conclusion
Quality control	Method Blank	< 0.20	< 0.2 mg/L	Pass
	Quality control Sample	19.3	17 to 23 mg/L	Pass
	Laboratory Fortified Blank	96.4	85 to 115 mg/L	Pass
	Duplicate Sample T1	3.3	% Difference within $\pm 15\%$	(%Diff = 4.2 %) Pass
	Duplicate Sample T2	3.4		

Note : 1. Perform method blank, quality control and laboratory fortified duplicate at a frequency of 1 in 20 samples or less.
2. mg/L denotes milligram per liter.
3. < denotes less than.
4. % denotes percent.

- End of report -

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park
18 On Lai Street, Shatin
NT, Hong Kong
Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>



Report No. : 01659
Application No. : HP01403

Issue Date : 15 Jul 2026

TEST REPORT

Applicant : Cinotech Consultants Limited
Rm 1710, Technology Park, 18 On Lai Street,
Shatin, New Territories, Hong Kong.

Sample Description : Four (4) submitted sample(s) stated to be Wastewater.

Project No. : MA25111

Sampling date : 8 Jul 2026

Sample Status Upon Receipt : Room Temperature

Date Received : 8 Jul 2026

Test Period : 9 Jul 2026 to 9 Jul 2026

Test Requested : Total Suspended Solids (TSS)

Test Method : Method SOP001 reference to Standard Methods for the Examination of Water and Wastewater, 23rd Ed, 2017- APHA, WWA, WPCF; Part 2540 D.

Test Result : Refer to the test result(s) on page 2.

Remark : 1. Information of the sample description (except sample status upon receipt) provided by the Applicant.
2. The result(s) relate only to the items tested.
3. The result(s) apply to the sample as received, where HPCT has not been responsible for the sampling stage.

For and on behalf of
HIGH PRECISION CHEMICAL TESTING LIMITED

Chan Hon Fai
Laboratory Director

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01659
Application No. : HP01403

Issue Date : 15 Jul 2026

TEST REPORT

Test Result : Total Suspended Solids (TSS)

Client sample ID	Sampling date	Laboratory ID	Result (mg/L)	Uncertainty (mg/L)
C1_1	8 Jul 2026	HP01403-01	4.7	± 0.3
C1_2	8 Jul 2026	HP01403-02	4.8	± 0.4
M1_1	8 Jul 2026	HP01403-03	3.1	± 0.2
M1_2	8 Jul 2026	HP01403-04	4.3	± 0.3

Note : 1. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%.
2. The reporting limit for Total Suspended Solids (TSS) is 1.0 mg/L.
3. mg/L denotes milligram per liter.

- End of report -

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park
18 On Lai Street, Shatin
NT, Hong Kong
Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>



Report No. : 01659QC
Application No. : HP01403

Issue Date : 10 Aug 2026

QC REPORT

Applicant : Cinotech Consultants Limited
Rm 1710, Technology Park, 18 On Lai Street,
Shatin, New Territories, Hong Kong.

Sample Description : Four (4) submitted sample(s) stated to be Wastewater.

Project No. : MA25111

Sampling date : 8 Jul 2026

Sample Status Upon Receipt : Room Temperature

Date Received : 8 Jul 2026

Test Period : 9 Jul 2026 to 9 Jul 2026

Test Requested : Total Suspended Solids (TSS)

Test Method : Method SOP001 reference to Standard Methods for the Examination of Water and Wastewater, 23rd Ed, 2017- APHA, WWA, WPCF; Part 2540 D.

QC Record : Refer to the record(s) on page 2.

Remark : Information of the sample description (except sample status upon receipt) provided by the Applicant.

For and on behalf of
HIGH PRECISION CHEMICAL TESTING LIMITED

Chan Hon Fai
Laboratory Director

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01659QC

Issue Date : 10 Aug 2026

Application No. : HP01403

QC REPORT

	Description	Reported Result (mg/L)	Acceptance criteria	Conclusion
Quality control	Method Blank	< 0.20	< 0.2 mg/L	Pass
	Quality control Sample	19.4	17 to 23 mg/L	Pass
	Laboratory Fortified Blank	96.1	85 to 115 mg/L	Pass
	Duplicate Sample T1	4.7	% Difference within $\pm 15\%$	(%Diff = 4.3 %) Pass
	Duplicate Sample T2	4.5		

Note : 1. Perform method blank, quality control and laboratory fortified duplicate at a frequency of 1 in 20 samples or less.
2. mg/L denotes milligram per liter.
3. < denotes less than.
4. % denotes percent.

- End of report -

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park
18 On Lai Street, Shatin
NT, Hong Kong
Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>



Report No. : 01660
Application No. : HP01404

Issue Date : 16 Jul 2026

TEST REPORT

Applicant : Cinotech Consultants Limited
Rm 1710, Technology Park, 18 On Lai Street,
Shatin, New Territories, Hong Kong.

Sample Description : Four (4) submitted sample(s) stated to be Wastewater.

Project No. : MA25111

Sampling date : 10 Jul 2026

Sample Status Upon Receipt : Room Temperature

Date Received : 10 Jul 2026

Test Period : 13 Jul 2026 to 13 Jul 2026

Test Requested : Total Suspended Solids (TSS)

Test Method : Method SOP001 reference to Standard Methods for the Examination of Water and Wastewater, 23rd Ed, 2017- APHA, WWA, WPCF; Part 2540 D.

Test Result : Refer to the test result(s) on page 2.

Remark : 1. Information of the sample description (except sample status upon receipt) provided by the Applicant.
2. The result(s) relate only to the items tested.
3. The result(s) apply to the sample as received, where HPCT has not been responsible for the sampling stage.

For and on behalf of
HIGH PRECISION CHEMICAL TESTING LIMITED

Chan Hon Fai
Laboratory Director

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01660
Application No. : HP01404

Issue Date : 16 Jul 2026

TEST REPORT

Test Result : Total Suspended Solids (TSS)

Client sample ID	Sampling date	Laboratory ID	Result (mg/L)	Uncertainty (mg/L)
C1_1	10 Jul 2026	HP01404-01	3.0	± 0.2
C1_2	10 Jul 2026	HP01404-02	2.8	± 0.2
M1_1	10 Jul 2026	HP01404-03	3.1	± 0.2
M1_2	10 Jul 2026	HP01404-04	3.7	± 0.3

Note : 1. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%.
2. The reporting limit for Total Suspended Solids (TSS) is 1.0 mg/L.
3. mg/L denotes milligram per liter.

- End of report -

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park
18 On Lai Street, Shatin
NT, Hong Kong
Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>



Report No. : 01660QC
Application No. : HP01404

Issue Date : 10 Aug 2026

QC REPORT

Applicant : Cinotech Consultants Limited
Rm 1710, Technology Park, 18 On Lai Street,
Shatin, New Territories, Hong Kong.

Sample Description : Four (4) submitted sample(s) stated to be Wastewater.

Project No. : MA25111

Sampling date : 10 Jul 2026

Sample Status Upon Receipt : Room Temperature

Date Received : 10 Jul 2026

Test Period : 13 Jul 2026 to 13 Jul 2026

Test Requested : Total Suspended Solids (TSS)

Test Method : Method SOP001 reference to Standard Methods for the Examination of Water and Wastewater, 23rd Ed, 2017- APHA, WWA, WPCF; Part 2540 D.

QC Record : Refer to the record(s) on page 2.

Remark : Information of the sample description (except sample status upon receipt) provided by the Applicant.

For and on behalf of
HIGH PRECISION CHEMICAL TESTING LIMITED

Chan Hon Fai
Laboratory Director

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01660QC

Issue Date : 10 Aug 2026

Application No. : HP01404

QC REPORT

	Description	Reported Result (mg/L)	Acceptance criteria	Conclusion
Quality control	Method Blank	< 0.20	< 0.2 mg/L	Pass
	Quality control Sample	19.2	17 to 23 mg/L	Pass
	Laboratory Fortified Blank	95.9	85 to 115 mg/L	Pass
	Duplicate Sample T1	3.0	% Difference within $\pm 15\%$	(%Diff = 6.5 %) Pass
	Duplicate Sample T2	3.2		

Note : 1. Perform method blank, quality control and laboratory fortified duplicate at a frequency of 1 in 20 samples or less.
2. mg/L denotes milligram per liter.
3. < denotes less than.
4. % denotes percent.

- End of report -

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park
18 On Lai Street, Shatin
NT, Hong Kong
Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>



Report No. : 01671
Application No. : HP01405

Issue Date : 20 Jul 2026

TEST REPORT

Applicant : Cinotech Consultants Limited
Rm 1710, Technology Park, 18 On Lai Street,
Shatin, New Territories, Hong Kong.

Sample Description : Four (4) submitted sample(s) stated to be Wastewater.

Project No. : MA25111

Sampling date : 13 Jul 2026

Sample Status Upon Receipt : Room Temperature

Date Received : 13 Jul 2026

Test Period : 14 Jul 2026 to 14 Jul 2026

Test Requested : Total Suspended Solids (TSS)

Test Method : Method SOP001 reference to Standard Methods for the Examination of Water and Wastewater, 23rd Ed, 2017- APHA, WWA, WPCF; Part 2540 D.

Test Result : Refer to the test result(s) on page 2.

Remark : 1. Information of the sample description (except sample status upon receipt) provided by the Applicant.
2. The result(s) relate only to the items tested.
3. The result(s) apply to the sample as received, where HPCT has not been responsible for the sampling stage.

For and on behalf of
HIGH PRECISION CHEMICAL TESTING LIMITED

Chan Hon Fai
Laboratory Director

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01671
Application No. : HP01405

Issue Date : 20 Jul 2026

TEST REPORT

Test Result : Total Suspended Solids (TSS)

Client sample ID	Sampling date	Laboratory ID	Result (mg/L)	Uncertainty (mg/L)
C1_1	13 Jul 2026	HP01405-01	21.5	± 1.6
C1_2	13 Jul 2026	HP01405-02	8.4	± 0.6
M1_1	13 Jul 2026	HP01405-03	9.3	± 0.7
M1_2	13 Jul 2026	HP01405-04	5.5	± 0.4

Note : 1. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%.
2. The reporting limit for Total Suspended Solids (TSS) is 1.0 mg/L.
3. mg/L denotes milligram per liter.

- End of report -

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park
18 On Lai Street, Shatin
NT, Hong Kong
Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>



Report No. : 01671QC
Application No. : HP01405

Issue Date : 10 Aug 2026

QC REPORT

Applicant : Cinotech Consultants Limited
Rm 1710, Technology Park, 18 On Lai Street,
Shatin, New Territories, Hong Kong.

Sample Description : Four (4) submitted sample(s) stated to be Wastewater.

Project No. : MA25111

Sampling date : 13 Jul 2026

Sample Status Upon Receipt : Room Temperature

Date Received : 13 Jul 2026

Test Period : 14 Jul 2026 to 14 Jul 2026

Test Requested : Total Suspended Solids (TSS)

Test Method : Method SOP001 reference to Standard Methods for the Examination of Water and Wastewater, 23rd Ed, 2017- APHA, WWA, WPCF; Part 2540 D.

QC Record : Refer to the record(s) on page 2.

Remark : Information of the sample description (except sample status upon receipt) provided by the Applicant.

For and on behalf of
HIGH PRECISION CHEMICAL TESTING LIMITED

Chan Hon Fai
Laboratory Director

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01671QC

Issue Date : 10 Aug 2026

Application No. : HP01405

QC REPORT

	Description	Reported Result (mg/L)	Acceptance criteria	Conclusion
Quality control	Method Blank	< 0.20	< 0.2 mg/L	Pass
	Quality control Sample	19.1	17 to 23 mg/L	Pass
	Laboratory Fortified Blank	95.7	85 to 115 mg/L	Pass
	Duplicate Sample T1	21.5	% Difference within $\pm 15\%$	(%Diff = 3.6 %) Pass
	Duplicate Sample T2	20.8		

Note : 1. Perform method blank, quality control and laboratory fortified duplicate at a frequency of 1 in 20 samples or less.
2. mg/L denotes milligram per liter.
3. < denotes less than.
4. % denotes percent.

- End of report -

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park
18 On Lai Street, Shatin
NT, Hong Kong
Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>



Report No. : 01672
Application No. : HP01406

Issue Date : 22 Jul 2026

TEST REPORT

Applicant : Cinotech Consultants Limited
Rm 1710, Technology Park, 18 On Lai Street,
Shatin, New Territories, Hong Kong.

Sample Description : Four (4) submitted sample(s) stated to be Wastewater.

Project No. : MA25111

Sampling date : 15 Jul 2026

Sample Status Upon Receipt : Room Temperature

Date Received : 15 Jul 2026

Test Period : 16 Jul 2026 to 16 Jul 2026

Test Requested : Total Suspended Solids (TSS)

Test Method : Method SOP001 reference to Standard Methods for the Examination of Water and Wastewater, 23rd Ed, 2017- APHA, WWA, WPCF; Part 2540 D.

Test Result : Refer to the test result(s) on page 2.

Remark : 1. Information of the sample description (except sample status upon receipt) provided by the Applicant.
2. The result(s) relate only to the items tested.
3. The result(s) apply to the sample as received, where HPCT has not been responsible for the sampling stage.

For and on behalf of
HIGH PRECISION CHEMICAL TESTING LIMITED

Chan Hon Fai
Laboratory Director

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01672
Application No. : HP01406

Issue Date : 22 Jul 2026

TEST REPORT

Test Result : Total Suspended Solids (TSS)

Client sample ID	Sampling date	Laboratory ID	Result (mg/L)	Uncertainty (mg/L)
C1_1	15 Jul 2026	HP01406-01	145	± 10.6
C1_2	15 Jul 2026	HP01406-02	152	± 11.1
M1_1	15 Jul 2026	HP01406-03	113	± 8.2
M1_2	15 Jul 2026	HP01406-04	109	± 7.9

Note : 1. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%.
2. The reporting limit for Total Suspended Solids (TSS) is 1.0 mg/L.
3. mg/L denotes milligram per liter.

- End of report -

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park
18 On Lai Street, Shatin
NT, Hong Kong
Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>



Report No. : 01672QC
Application No. : HP01406

Issue Date : 10 Aug 2026

QC REPORT

Applicant : Cinotech Consultants Limited
Rm 1710, Technology Park, 18 On Lai Street,
Shatin, New Territories, Hong Kong.

Sample Description : Four (4) submitted sample(s) stated to be Wastewater.

Project No. : MA25111

Sampling date : 15 Jul 2026

Sample Status Upon Receipt : Room Temperature

Date Received : 15 Jul 2026

Test Period : 16 Jul 2026 to 16 Jul 2026

Test Requested : Total Suspended Solids (TSS)

Test Method : Method SOP001 reference to Standard Methods for the Examination of Water and Wastewater, 23rd Ed, 2017- APHA, WWA, WPCF; Part 2540 D.

QC Record : Refer to the record(s) on page 2.

Remark : Information of the sample description (except sample status upon receipt) provided by the Applicant.

For and on behalf of
HIGH PRECISION CHEMICAL TESTING LIMITED

Chan Hon Fai
Laboratory Director

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01672QC

Issue Date : 10 Aug 2026

Application No. : HP01406

QC REPORT

	Description	Reported Result (mg/L)	Acceptance criteria	Conclusion
Quality control	Method Blank	< 0.20	< 0.2 mg/L	Pass
	Quality control Sample	19.5	17 to 23 mg/L	Pass
	Laboratory Fortified Blank	96.9	85 to 115 mg/L	Pass
	Duplicate Sample T1	145	% Difference within $\pm 15\%$	(%Diff = 3.5 %) Pass
	Duplicate Sample T2	150		

Note : 1. Perform method blank, quality control and laboratory fortified duplicate at a frequency of 1 in 20 samples or less.
2. mg/L denotes milligram per liter.
3. < denotes less than.
4. % denotes percent.

- End of report -

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park
18 On Lai Street, Shatin
NT, Hong Kong
Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>



Report No. : 01673
Application No. : HP01407

Issue Date : 24 Jul 2026

TEST REPORT

Applicant : Cinotech Consultants Limited
Rm 1710, Technology Park, 18 On Lai Street,
Shatin, New Territories, Hong Kong.

Sample Description : Four (4) submitted sample(s) stated to be Wastewater.

Project No. : MA25111

Sampling date : 17 Jul 2026

Sample Status Upon Receipt : Room Temperature

Date Received : 17 Jul 2026

Test Period : 20 Jul 2026 to 20 Jul 2026

Test Requested : Total Suspended Solids (TSS)

Test Method : Method SOP001 reference to Standard Methods for the Examination of Water and Wastewater, 23rd Ed, 2017- APHA, WWA, WPCF; Part 2540 D.

Test Result : Refer to the test result(s) on page 2.

Remark : 1. Information of the sample description (except sample status upon receipt) provided by the Applicant.
2. The result(s) relate only to the items tested.
3. The result(s) apply to the sample as received, where HPCT has not been responsible for the sampling stage.

For and on behalf of
HIGH PRECISION CHEMICAL TESTING LIMITED

Chan Hon Fai
Laboratory Director

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01673
Application No. : HP01407

Issue Date : 24 Jul 2026

TEST REPORT

Test Result : Total Suspended Solids (TSS)

Client sample ID	Sampling date	Laboratory ID	Result (mg/L)	Uncertainty (mg/L)
C1_1	17 Jul 2026	HP01407-01	38.5	± 2.8
C1_2	17 Jul 2026	HP01407-02	45.9	± 3.3
M1_1	17 Jul 2026	HP01407-03	139	± 10.2
M1_2	17 Jul 2026	HP01407-04	133	± 9.7

Note : 1. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%.
2. The reporting limit for Total Suspended Solids (TSS) is 1.0 mg/L.
3. mg/L denotes milligram per liter.

- End of report -

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park
18 On Lai Street, Shatin
NT, Hong Kong
Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>



Report No. : 01673QC
Application No. : HP01407

Issue Date : 10 Aug 2026

QC REPORT

Applicant : Cinotech Consultants Limited
Rm 1710, Technology Park, 18 On Lai Street,
Shatin, New Territories, Hong Kong.

Sample Description : Four (4) submitted sample(s) stated to be Wastewater.

Project No. : MA25111

Sampling date : 17 Jul 2026

Sample Status Upon Receipt : Room Temperature

Date Received : 17 Jul 2026

Test Period : 20 Jul 2026 to 20 Jul 2026

Test Requested : Total Suspended Solids (TSS)

Test Method : Method SOP001 reference to Standard Methods for the Examination of Water and Wastewater, 23rd Ed, 2017- APHA, WWA, WPCF; Part 2540 D.

QC Record : Refer to the record(s) on page 2.

Remark : Information of the sample description (except sample status upon receipt) provided by the Applicant.

For and on behalf of
HIGH PRECISION CHEMICAL TESTING LIMITED

Chan Hon Fai
Laboratory Director

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01673QC

Issue Date : 10 Aug 2026

Application No. : HP01407

QC REPORT

	Description	Reported Result (mg/L)	Acceptance criteria	Conclusion
Quality control	Method Blank	< 0.20	< 0.2 mg/L	Pass
	Quality control Sample	19.1	17 to 23 mg/L	Pass
	Laboratory Fortified Blank	95.2	85 to 115 mg/L	Pass
	Duplicate Sample T1	38.5	% Difference within $\pm 15\%$	(%Diff = 0.8 %) Pass
	Duplicate Sample T2	38.8		

Note : 1. Perform method blank, quality control and laboratory fortified duplicate at a frequency of 1 in 20 samples or less.
2. mg/L denotes milligram per liter.
3. < denotes less than.
4. % denotes percent.

- End of report -

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park
18 On Lai Street, Shatin
NT, Hong Kong
Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>



Report No. : 01686
Application No. : HP01426

Issue Date : 27 Jul 2026

TEST REPORT

Applicant : Cinotech Consultants Limited
Rm 1710, Technology Park, 18 On Lai Street,
Shatin, New Territories, Hong Kong.

Sample Description : Four (4) submitted sample(s) stated to be Wastewater.

Project No. : MA25111

Sampling date : 20 Jul 2026

Sample Status Upon Receipt : Room Temperature

Date Received : 20 Jul 2026

Test Period : 21 Jul 2026 to 21 Jul 2026

Test Requested : Total Suspended Solids (TSS)

Test Method : Method SOP001 reference to Standard Methods for the Examination of Water and Wastewater, 23rd Ed, 2017- APHA, WWA, WPCF; Part 2540 D.

Test Result : Refer to the test result(s) on page 2.

Remark : 1. Information of the sample description (except sample status upon receipt) provided by the Applicant.
2. The result(s) relate only to the items tested.
3. The result(s) apply to the sample as received, where HPCT has not been responsible for the sampling stage.

For and on behalf of
HIGH PRECISION CHEMICAL TESTING LIMITED

Chan Hon Fai
Laboratory Director

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01686
Application No. : HP01426

Issue Date : 27 Jul 2026

TEST REPORT

Test Result : Total Suspended Solids (TSS)

Client sample ID	Sampling date	Laboratory ID	Result (mg/L)	Uncertainty (mg/L)
C1_1	20 Jul 2026	HP01426-01	19.1	± 1.4
C1_2	20 Jul 2026	HP01426-02	19.3	± 1.4
M1_1	20 Jul 2026	HP01426-03	66.8	± 4.9
M1_2	20 Jul 2026	HP01426-04	68.2	± 5.0

Note : 1. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%.
2. The reporting limit for Total Suspended Solids (TSS) is 1.0 mg/L.
3. mg/L denotes milligram per liter.

- End of report -

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park
18 On Lai Street, Shatin
NT, Hong Kong
Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>



Report No. : 01686QC
Application No. : HP01426

Issue Date : 10 Aug 2026

QC REPORT

Applicant : Cinotech Consultants Limited
Rm 1710, Technology Park, 18 On Lai Street,
Shatin, New Territories, Hong Kong.

Sample Description : Four (4) submitted sample(s) stated to be Wastewater.

Project No. : MA25111

Sampling date : 20 Jul 2026

Sample Status Upon Receipt : Room Temperature

Date Received : 20 Jul 2026

Test Period : 21 Jul 2026 to 21 Jul 2026

Test Requested : Total Suspended Solids (TSS)

Test Method : Method SOP001 reference to Standard Methods for the Examination of Water and Wastewater, 23rd Ed, 2017- APHA, WWA, WPCF; Part 2540 D.

QC Record : Refer to the record(s) on page 2.

Remark : Information of the sample description (except sample status upon receipt) provided by the Applicant.

For and on behalf of
HIGH PRECISION CHEMICAL TESTING LIMITED

Chan Hon Fai
Laboratory Director

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01686QC

Issue Date : 10 Aug 2026

Application No. : HP01426

QC REPORT

	Description	Reported Result (mg/L)	Acceptance criteria	Conclusion
Quality control	Method Blank	< 0.20	< 0.2 mg/L	Pass
	Quality control Sample	20.3	17 to 23 mg/L	Pass
	Laboratory Fortified Blank	98.6	85 to 115 mg/L	Pass
	Duplicate Sample T1	19.1	% Difference within $\pm 15\%$	(%Diff = 4.8 %) Pass
	Duplicate Sample T2	20.0		

Note : 1. Perform method blank, quality control and laboratory fortified duplicate at a frequency of 1 in 20 samples or less.
2. mg/L denotes milligram per liter.
3. < denotes less than.
4. % denotes percent.

- End of report -

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park
18 On Lai Street, Shatin
NT, Hong Kong
Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>



Report No. : 01687
Application No. : HP01427

Issue Date : 29 Jul 2026

TEST REPORT

Applicant : Cinotech Consultants Limited
Rm 1710, Technology Park, 18 On Lai Street,
Shatin, New Territories, Hong Kong.

Sample Description : Four (4) submitted sample(s) stated to be Wastewater.

Project No. : MA25111

Sampling date : 22 Jul 2026

Sample Status Upon Receipt : Room Temperature

Date Received : 22 Jul 2026

Test Period : 23 Jul 2026 to 23 Jul 2026

Test Requested : Total Suspended Solids (TSS)

Test Method : Method SOP001 reference to Standard Methods for the Examination of Water and Wastewater, 23rd Ed, 2017- APHA, WWA, WPCF; Part 2540 D.

Test Result : Refer to the test result(s) on page 2.

Remark : 1. Information of the sample description (except sample status upon receipt) provided by the Applicant.
2. The result(s) relate only to the items tested.
3. The result(s) apply to the sample as received, where HPCT has not been responsible for the sampling stage.

For and on behalf of
HIGH PRECISION CHEMICAL TESTING LIMITED

Chan Hon Fai
Laboratory Director

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01687
Application No. : HP01427

Issue Date : 29 Jul 2026

TEST REPORT

Test Result : Total Suspended Solids (TSS)

Client sample ID	Sampling date	Laboratory ID	Result (mg/L)	Uncertainty (mg/L)
C1_1	22 Jul 2026	HP01427-01	4.4	± 0.3
C1_2	22 Jul 2026	HP01427-02	4.7	± 0.3
M1_1	22 Jul 2026	HP01427-03	5.0	± 0.4
M1_2	22 Jul 2026	HP01427-04	4.3	± 0.3

Note : 1. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%.
2. The reporting limit for Total Suspended Solids (TSS) is 1.0 mg/L.
3. mg/L denotes milligram per liter.

- End of report -

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park
18 On Lai Street, Shatin
NT, Hong Kong
Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>



Report No. : 01687QC
Application No. : HP01427

Issue Date : 10 Aug 2026

QC REPORT

Applicant : Cinotech Consultants Limited
Rm 1710, Technology Park, 18 On Lai Street,
Shatin, New Territories, Hong Kong.

Sample Description : Four (4) submitted sample(s) stated to be Wastewater.

Project No. : MA25111

Sampling date : 22 Jul 2026

Sample Status Upon Receipt : Room Temperature

Date Received : 22 Jul 2026

Test Period : 23 Jul 2026 to 23 Jul 2026

Test Requested : Total Suspended Solids (TSS)

Test Method : Method SOP001 reference to Standard Methods for the Examination of Water and Wastewater, 23rd Ed, 2017- APHA, WWA, WPCF; Part 2540 D.

QC Record : Refer to the record(s) on page 2.

Remark : Information of the sample description (except sample status upon receipt) provided by the Applicant.

For and on behalf of
HIGH PRECISION CHEMICAL TESTING LIMITED

Chan Hon Fai
Laboratory Director

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01687QC

Issue Date : 10 Aug 2026

Application No. : HP01427

QC REPORT

	Description	Reported Result (mg/L)	Acceptance criteria	Conclusion
Quality control	Method Blank	< 0.20	< 0.2 mg/L	Pass
	Quality control Sample	19.3	17 to 23 mg/L	Pass
	Laboratory Fortified Blank	96.0	85 to 115 mg/L	Pass
	Duplicate Sample T1	4.4	% Difference within $\pm 15\%$	(%Diff = 1.3 %) Pass
	Duplicate Sample T2	4.5		

Note : 1. Perform method blank, quality control and laboratory fortified duplicate at a frequency of 1 in 20 samples or less.
2. mg/L denotes milligram per liter.
3. < denotes less than.
4. % denotes percent.

- End of report -

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park
18 On Lai Street, Shatin
NT, Hong Kong
Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>



Report No. : 01688
Application No. : HP01428

Issue Date : 31 Jul 2026

TEST REPORT

Applicant : Cinotech Consultants Limited
Rm 1710, Technology Park, 18 On Lai Street,
Shatin, New Territories, Hong Kong.

Sample Description : Four (4) submitted sample(s) stated to be Wastewater.

Project No. : MA25111

Sampling date : 24 Jul 2026

Sample Status Upon Receipt : Room Temperature

Date Received : 24 Jul 2026

Test Period : 27 Jul 2026 to 27 Jul 2026

Test Requested : Total Suspended Solids (TSS)

Test Method : Method SOP001 reference to Standard Methods for the Examination of Water and Wastewater, 23rd Ed, 2017- APHA, WWA, WPCF; Part 2540 D.

Test Result : Refer to the test result(s) on page 2.

Remark : 1. Information of the sample description (except sample status upon receipt) provided by the Applicant.
2. The result(s) relate only to the items tested.
3. The result(s) apply to the sample as received, where HPCT has not been responsible for the sampling stage.

For and on behalf of
HIGH PRECISION CHEMICAL TESTING LIMITED

Chan Hon Fai
Laboratory Director

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01688
Application No. : HP01428

Issue Date : 31 Jul 2026

TEST REPORT

Test Result : Total Suspended Solids (TSS)

Client sample ID	Sampling date	Laboratory ID	Result (mg/L)	Uncertainty (mg/L)
C1_1	24 Jul 2026	HP01428-01	3.7	± 0.3
C1_2	24 Jul 2026	HP01428-02	3.5	± 0.3
M1_1	24 Jul 2026	HP01428-03	2.6	± 0.2
M1_2	24 Jul 2026	HP01428-04	2.3	± 0.2

Note : 1. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%.
2. The reporting limit for Total Suspended Solids (TSS) is 1.0 mg/L.
3. mg/L denotes milligram per liter.

- End of report -

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01688QC

Issue Date : 10 Aug 2026

Application No. : HP01428

QC REPORT

Applicant : Cinotech Consultants Limited
Rm 1710, Technology Park, 18 On Lai Street,
Shatin, New Territories, Hong Kong.

Sample Description : Four (4) submitted sample(s) stated to be Wastewater.

Project No. : MA25111

Sampling date : 24 Jul 2026

Sample Status Upon Receipt : Room Temperature

Date Received : 24 Jul 2026

Test Period : 27 Jul 2026 to 27 Jul 2026

Test Requested : Total Suspended Solids (TSS)

Test Method : Method SOP001 reference to Standard Methods for the Examination of Water and Wastewater, 23rd Ed, 2017- APHA, WWA, WPCF; Part 2540 D.

QC Record : Refer to the record(s) on page 2.

Remark : Information of the sample description (except sample status upon receipt) provided by the Applicant.

For and on behalf of
HIGH PRECISION CHEMICAL TESTING LIMITED

Chan Hon Fai
Laboratory Director

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01688QC

Issue Date : 10 Aug 2026

Application No. : HP01428

QC REPORT

	Description	Reported Result (mg/L)	Acceptance criteria	Conclusion
Quality control	Method Blank	< 0.20	< 0.2 mg/L	Pass
	Quality control Sample	19.9	17 to 23 mg/L	Pass
	Laboratory Fortified Blank	96.7	85 to 115 mg/L	Pass
	Duplicate Sample T1	3.7	% Difference within $\pm 15\%$	(%Diff = 2.2 %) Pass
	Duplicate Sample T2	3.8		

Note : 1. Perform method blank, quality control and laboratory fortified duplicate at a frequency of 1 in 20 samples or less.
2. mg/L denotes milligram per liter.
3. < denotes less than.
4. % denotes percent.

- End of report -

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park
18 On Lai Street, Shatin
NT, Hong Kong
Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>



Report No. : 01698
Application No. : HP01429

Issue Date : 3 Aug 2026

TEST REPORT

Applicant : Cinotech Consultants Limited
Rm 1710, Technology Park, 18 On Lai Street,
Shatin, New Territories, Hong Kong.

Sample Description : Four (4) submitted sample(s) stated to be Wastewater.

Project No. : MA25111

Sampling date : 27 Jul 2026

Sample Status Upon Receipt : Room Temperature

Date Received : 27 Jul 2026

Test Period : 28 Jul 2026 to 28 Jul 2026

Test Requested : Total Suspended Solids (TSS)

Test Method : Method SOP001 reference to Standard Methods for the Examination of Water and Wastewater, 23rd Ed, 2017- APHA, WWA, WPCF; Part 2540 D.

Test Result : Refer to the test result(s) on page 2.

Remark : 1. Information of the sample description (except sample status upon receipt) provided by the Applicant.
2. The result(s) relate only to the items tested.
3. The result(s) apply to the sample as received, where HPCT has not been responsible for the sampling stage.

For and on behalf of
HIGH PRECISION CHEMICAL TESTING LIMITED

Chan Hon Fai
Laboratory Director

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01698
Application No. : HP01429

Issue Date : 3 Aug 2026

TEST REPORT

Test Result : Total Suspended Solids (TSS)

Client sample ID	Sampling date	Laboratory ID	Result (mg/L)	Uncertainty (mg/L)
C1_1	27 Jul 2026	HP01429-01	9.0	± 0.7
C1_2	27 Jul 2026	HP01429-02	9.1	± 0.7
M1_1	27 Jul 2026	HP01429-03	19.5	± 1.4
M1_2	27 Jul 2026	HP01429-04	19.7	± 1.4

Note : 1. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%.
2. The reporting limit for Total Suspended Solids (TSS) is 1.0 mg/L.
3. mg/L denotes milligram per liter.

- End of report -

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park
18 On Lai Street, Shatin
NT, Hong Kong
Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>



Report No. : 01698QC
Application No. : HP01429

Issue Date : 10 Aug 2026

QC REPORT

Applicant : Cinotech Consultants Limited
Rm 1710, Technology Park, 18 On Lai Street,
Shatin, New Territories, Hong Kong.

Sample Description : Four (4) submitted sample(s) stated to be Wastewater.

Project No. : MA25111

Sampling date : 27 Jul 2026

Sample Status Upon Receipt : Room Temperature

Date Received : 27 Jul 2026

Test Period : 28 Jul 2026 to 28 Jul 2026

Test Requested : Total Suspended Solids (TSS)

Test Method : Method SOP001 reference to Standard Methods for the Examination of Water and Wastewater, 23rd Ed, 2017- APHA, WWA, WPCF; Part 2540 D.

QC Record : Refer to the record(s) on page 2.

Remark : Information of the sample description (except sample status upon receipt) provided by the Applicant.

For and on behalf of
HIGH PRECISION CHEMICAL TESTING LIMITED

Chan Hon Fai
Laboratory Director

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01698QC

Issue Date : 10 Aug 2026

Application No. : HP01429

QC REPORT

	Description	Reported Result (mg/L)	Acceptance criteria	Conclusion
Quality control	Method Blank	< 0.20	< 0.2 mg/L	Pass
	Quality control Sample	19.3	17 to 23 mg/L	Pass
	Laboratory Fortified Blank	97.2	85 to 115 mg/L	Pass
	Duplicate Sample T1	9.0	% Difference within $\pm 15\%$	(%Diff = 2.6 %) Pass
	Duplicate Sample T2	9.2		

Note : 1. Perform method blank, quality control and laboratory fortified duplicate at a frequency of 1 in 20 samples or less.
2. mg/L denotes milligram per liter.
3. < denotes less than.
4. % denotes percent.

- End of report -

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park
18 On Lai Street, Shatin
NT, Hong Kong
Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>



Report No. : 01699
Application No. : HP01430

Issue Date : 5 Aug 2026

TEST REPORT

Applicant : Cinotech Consultants Limited
Rm 1710, Technology Park, 18 On Lai Street,
Shatin, New Territories, Hong Kong.

Sample Description : Four (4) submitted sample(s) stated to be Wastewater.

Project No. : MA25111

Sampling date : 29 Jul 2026

Sample Status Upon Receipt : Room Temperature

Date Received : 29 Jul 2026

Test Period : 30 Jul 2026 to 30 Jul 2026

Test Requested : Total Suspended Solids (TSS)

Test Method : Method SOP001 reference to Standard Methods for the Examination of Water and Wastewater, 23rd Ed, 2017- APHA, WWA, WPCF; Part 2540 D.

Test Result : Refer to the test result(s) on page 2.

Remark : 1. Information of the sample description (except sample status upon receipt) provided by the Applicant.
2. The result(s) relate only to the items tested.
3. The result(s) apply to the sample as received, where HPCT has not been responsible for the sampling stage.

For and on behalf of
HIGH PRECISION CHEMICAL TESTING LIMITED

Chan Hon Fai
Laboratory Director

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01699
Application No. : HP01430

Issue Date : 5 Aug 2026

TEST REPORT

Test Result : Total Suspended Solids (TSS)

Client sample ID	Sampling date	Laboratory ID	Result (mg/L)	Uncertainty (mg/L)
C1_1	29 Jul 2026	HP01430-01	11.9	± 0.9
C1_2	29 Jul 2026	HP01430-02	11.8	± 0.9
M1_1	29 Jul 2026	HP01430-03	28.3	± 2.1
M1_2	29 Jul 2026	HP01430-04	13.1	± 1.0

Note : 1. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%.
2. The reporting limit for Total Suspended Solids (TSS) is 1.0 mg/L.
3. mg/L denotes milligram per liter.

- End of report -

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park
18 On Lai Street, Shatin
NT, Hong Kong
Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>



Report No. : 01699QC
Application No. : HP01430

Issue Date : 10 Aug 2026

QC REPORT

Applicant : Cinotech Consultants Limited
Rm 1710, Technology Park, 18 On Lai Street,
Shatin, New Territories, Hong Kong.

Sample Description : Four (4) submitted sample(s) stated to be Wastewater.

Project No. : MA25111

Sampling date : 29 Jul 2026

Sample Status Upon Receipt : Room Temperature

Date Received : 29 Jul 2026

Test Period : 30 Jul 2026 to 30 Jul 2026

Test Requested : Total Suspended Solids (TSS)

Test Method : Method SOP001 reference to Standard Methods for the Examination of Water and Wastewater, 23rd Ed, 2017- APHA, WWA, WPCF; Part 2540 D.

QC Record : Refer to the record(s) on page 2.

Remark : Information of the sample description (except sample status upon receipt) provided by the Applicant.

For and on behalf of
HIGH PRECISION CHEMICAL TESTING LIMITED

Chan Hon Fai
Laboratory Director

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01699QC

Issue Date : 10 Aug 2026

Application No. : HP01430

QC REPORT

	Description	Reported Result (mg/L)	Acceptance criteria	Conclusion
Quality control	Method Blank	< 0.20	< 0.2 mg/L	Pass
	Quality control Sample	19.5	17 to 23 mg/L	Pass
	Laboratory Fortified Blank	96.5	85 to 115 mg/L	Pass
	Duplicate Sample T1	11.9	% Difference within $\pm 15\%$	(%Diff = 4.4 %) Pass
	Duplicate Sample T2	12.5		

Note : 1. Perform method blank, quality control and laboratory fortified duplicate at a frequency of 1 in 20 samples or less.
2. mg/L denotes milligram per liter.
3. < denotes less than.
4. % denotes percent.

- End of report -

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park
18 On Lai Street, Shatin
NT, Hong Kong
Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>



Report No. : 01700
Application No. : HP01431

Issue Date : 7 Aug 2026

TEST REPORT

Applicant : Cinotech Consultants Limited
Rm 1710, Technology Park, 18 On Lai Street,
Shatin, New Territories, Hong Kong.

Sample Description : Four (4) submitted sample(s) stated to be Wastewater.

Project No. : MA25111

Sampling date : 31 Jul 2026

Sample Status Upon Receipt : Room Temperature

Date Received : 31 Jul 2026

Test Period : 3 Aug 2026 to 3 Aug 2026

Test Requested : Total Suspended Solids (TSS)

Test Method : Method SOP001 reference to Standard Methods for the Examination of Water and Wastewater, 23rd Ed, 2017- APHA, WWA, WPCF; Part 2540 D.

Test Result : Refer to the test result(s) on page 2.

Remark : 1. Information of the sample description (except sample status upon receipt) provided by the Applicant.
2. The result(s) relate only to the items tested.
3. The result(s) apply to the sample as received, where HPCT has not been responsible for the sampling stage.

For and on behalf of
HIGH PRECISION CHEMICAL TESTING LIMITED

Chan Hon Fai
Laboratory Director

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01700
Application No. : HP01431

Issue Date : 7 Aug 2026

TEST REPORT

Test Result : Total Suspended Solids (TSS)

Client sample ID	Sampling date	Laboratory ID	Result (mg/L)	Uncertainty (mg/L)
C1_1	31 Jul 2026	HP01431-01	35.8	± 2.6
C1_2	31 Jul 2026	HP01431-02	34.4	± 2.5
M1_1	31 Jul 2026	HP01431-03	97.6	± 7.1
M1_2	31 Jul 2026	HP01431-04	27.6	± 2.0

Note : 1. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%.
2. The reporting limit for Total Suspended Solids (TSS) is 1.0 mg/L.
3. mg/L denotes milligram per liter.

- End of report -

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01700QC

Issue Date : 10 Aug 2026

Application No. : HP01431

QC REPORT

Applicant : Cinotech Consultants Limited
Rm 1710, Technology Park, 18 On Lai Street,
Shatin, New Territories, Hong Kong.

Sample Description : Four (4) submitted sample(s) stated to be Wastewater.

Project No. : MA25111

Sampling date : 31 Jul 2026

Sample Status Upon Receipt : Room Temperature

Date Received : 31 Jul 2026

Test Period : 3 Aug 2026 to 3 Aug 2026

Test Requested : Total Suspended Solids (TSS)

Test Method : Method SOP001 reference to Standard Methods for the Examination of Water and Wastewater, 23rd Ed, 2017- APHA, WWA, WPCF; Part 2540 D.

QC Record : Refer to the record(s) on page 2.

Remark : Information of the sample description (except sample status upon receipt) provided by the Applicant.

For and on behalf of
HIGH PRECISION CHEMICAL TESTING LIMITED

Chan Hon Fai
Laboratory Director

High Precision Chemical Testing Ltd.

Rm 1904, Technology Park

18 On Lai Street, Shatin

NT, Hong Kong

Tel: +852 3841 4388 Website: <https://www.hpct.com.hk>

Report No. : 01700QC

Issue Date : 10 Aug 2026

Application No. : HP01431

QC REPORT

	Description	Reported Result (mg/L)	Acceptance criteria	Conclusion
Quality control	Method Blank	< 0.20	< 0.2 mg/L	Pass
	Quality control Sample	18.9	17 to 23 mg/L	Pass
	Laboratory Fortified Blank	95.9	85 to 115 mg/L	Pass
	Duplicate Sample T1	35.8	% Difference within $\pm 15\%$	(%Diff = 5.0 %) Pass
	Duplicate Sample T2	37.7		

- Note** : 1. Perform method blank, quality control and laboratory fortified duplicate at a frequency of 1 in 20 samples or less.
2. mg/L denotes milligram per liter.
3. < denotes less than.
4. % denotes percent.

- End of report -